Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX: 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0138 | Reference Number: | 83-124312941-1 |
| Cylinder Number: | CC14644 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B42012 | Valve Outlet: | 350 |
| Gas Code: | APPVD | Analysis Date: | Apr 27, 2012 |

**Expiration Date: Apr 27, 2015**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| CARBON MONOXIDE | 2500 PPM | 2472 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 00052503 | SG9160725BAL | 1985PPM CARBON MONOXIDE/NITROGEN | Apr 15, 2012 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| DH3CO | NonDispersive Infrared | Apr 02 2012 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR Audit

Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

**Airgas Specialty Gases**
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX: 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0138 | Reference Number: | 83-124312941-1 |
| Cylinder Number: | CC22326 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B42012 | Valve Outlet: | 350 |
| Gas Code: | APPVD | Analysis Date: | Apr 27, 2012 |

**Expiration Date: Apr 27, 2015**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| CARBON MONOXIDE | 2500 PPM | 2480 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 00052503 | SG9160725BAL | 1985PPM CARBON MONOXIDE/NITROGEN | Apr 15, 2012 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| DH3CO | NonDispersive Infrared | Apr 02, 2012 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR Prec/CAL

# Airgas

**Airgas Specialty Gases**
12722 South Wentworth Avenue
Chicago, IL 60628
(773) 785-3000 Fax: (773) 785-1928
www.airgas.com

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0083 | Reference Number: | 54-124341745-1 |
| Cylinder Number: | CC213280 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Chicago - IL | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B12012 | Valve Outlet: | 350 |
| Gas Code: | APPVD | Analysis Date: | Oct 20, 2012 |

**Expiration Date: Oct 20, 2020**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| CARBON MONOXIDE | 170.0 PPM | 167.4 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM/CO | 09060434 | CC287237 | 501.3PPM CARBON MONOXIDE/ | Feb 01, 2013 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| CO-1 HORIBA VIA-510 TKPPF7FG | NDIR | Oct 15, 2012 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR Audit

SK

# EPA Protocol Gas Mixture

**Certified Concentrations**

| Component | CASNumber | Concentration | Accuracy | Procedure |
|---|---|---|---|---|
| CARBON MONOXIDE | 630-08-0 | 171.4 PPM | +/-1% | G1 |
| NITROGEN | 7727-37-9 | Balance | | |

**Cylinder Number:** CC488846
Cylinder Pressure: 2015 PSIG
Certification Date: Oct 20, 2012
Expiration Date: Oct 20, 2020
Reference Number: 54-124341745-1
Part Number: E02NI99E15A0083
PGVP Number: B12012
Gas Code: APPVD

**Notes:**
Do not use cylinder below 100 psig.
Certification performed in accordance with "EPA Traceability Protocol document EPA 600/R-12/531 (May 2012)" using assay procedures listed.

To reorder this mixture, use Part Number: E02NI99E15A0083

12722 S. Wentworth Ave. . B1 Chicago IL 60628

PARR Precision / Calibration

Airgas

Airgas Specialty Gases
630 United Drive
Durham, NC 27713
919-544-3773 Fax: 919-544-3774
www.airgas.com

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0383 | Reference Number: | 122-124342451-1 |
| Cylinder Number: | CC421756 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Durham - NC | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B22012 | Valve Outlet: | 660 |
| Gas Code: | SO2 | Analysis Date: | Nov 05, 2012 |

**Expiration Date: Nov 05, 2016**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 15.00 PPM | 15.22 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 100610 | CC284538 | 14.82PPM SULFUR DIOXIDE/NITROGEN | Jul 13, 2013 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801333 SO2 | FTIR | Oct 11, 2012 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

Audit Cylinder

SD

Airgas

Airgas Specialty Gases
630 United Drive
Durham, NC 27713
919-544-3773 Fax: 919-544-3774
www.airgas.com

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0383 | Reference Number: | 122-124342451-1 |
| Cylinder Number: | CC421981 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Durham - NC | Cylinder Pressure: | 2015 PSIG |
| PGVP Number: | B22012 | Valve Outlet: | 660 |
| Gas Code: | SO2 | Analysis Date: | Nov 05, 2012 |

**Expiration Date: Nov 05, 2016**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 15.00 PPM | 15.13 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 100610 | CC284538 | 14.82PPM SULFUR DIOXIDE/NITROGEN | Jul 13, 2013 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801333 SO2 | FTIR | Oct 11, 2012 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR Calibration/Precision

# Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0350 | Reference Number: | 83-124256035-2 |
| Cylinder Number: | CC343761 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| Analysis Date: | Mar 15, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 15, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 50.00 PPM | 48.30 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 08061508 | CC254794 | 94.67PPM SULFUR DIOXIDE/NITROGEN | Oct 15, 2012 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LSO2 | FTIR | Feb 18, 2011 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

LAB

# Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E80A0202 | Reference Number: | 83-124256035-3 |
| Cylinder Number: | LL7671 | Cylinder Volume: | 83 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2215 PSIG |
| Analysis Date: | Mar 15, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 15, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 50.00 PPM | 49.62 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 08061508 | CC254794 | 94.67PPM SULFUR DIOXIDE/NITROGEN | Oct 15, 2012 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LSO2 | FTIR | Feb 18, 2011 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

Lab Audit

SR

# Airgas

# CERTIFICATE OF ANALYSIS
# Grade of Product: EPA Protocol

**Airgas Specialty Gases**
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E80A0288 | Reference Number: | 83-124256035-4A |
| Cylinder Number: | SG9115506ALC | Cylinder Volume: | 83 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2215 PSIG |
| Analysis Date: | Mar 22, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 22, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

## ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| NITRIC OXIDE | 50.00 PPM | 50.61 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |
| Total oxides of nitrogen | | 50.80 PPM | | For Reference Only |

## CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 10061119 | cc283757 | 49.73PPM NITRIC OXIDE/NITROGEN | Jul 23, 2016 |

## ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LNO | FTIR | Feb 28, 2011 |

**Triad Data Available Upon Request**

Notes:

Approved for Release

Lab Audit

# Airgas

# CERTIFICATE OF ANALYSIS
# Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0350 | Reference Number: | 83-124256035-2 |
| Cylinder Number: | CC20611 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| Analysis Date: | Mar 15, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 15, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

## ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 50.00 PPM | 49.00 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

## CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 08061508 | CC254794 | 94.67PPM SULFUR DIOXIDE/NITROGEN | Oct 15, 2012 |

## ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LSO2 | FTIR | Feb 18, 2011 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

CAL
ELDO $SO_2$ Precision

SR

**Airgas**

# CERTIFICATE OF ANALYSIS
# Grade of Product: EPA Protocol

**Airgas Specialty Gases**
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0350 | Reference Number: | 83-124256035-2 |
| Cylinder Number: | CC334092 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| Analysis Date: | Mar 15, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 15, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

## ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| SULFUR DIOXIDE | 50.00 PPM | 49.04 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

## CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 08061508 | CC254794 | 94.67PPM SULFUR DIOXIDE/NITROGEN | Oct 15, 2012 |

## ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LSO2 | FTIR | Feb 18, 2011 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR

CAL/Precision

# Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0147 | Reference Number: | 83-124256035-1 |
| Cylinder Number: | CC208157 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| Analysis Date: | Mar 14, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 14, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| NITRIC OXIDE | 50.00 PPM | 51.78 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |
| Total oxides of nitrogen | | 51.81 PPM | | For Reference Only |

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 10061119 | cc283757 | 49.73PPM NITRIC OXIDE/NITROGEN | Jul 23, 2016 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LNO | FTIR | Feb 28, 2011 |

**Triad Data Available Upon Request**

Notes:

Approved for Release

Marion

CAL/Precision

# Airgas

# CERTIFICATE OF ANALYSIS
## Grade of Product: EPA Protocol

Airgas Specialty Gases
1075 Cinclare Drive
Port Allen, LA 70767
225.388.0900
FAX : 225.388.0959
www.airgas.com

| | | | |
|---|---|---|---|
| Part Number: | E02NI99E15A0147 | Reference Number: | 83-124256035-1 |
| Cylinder Number: | CC221153 | Cylinder Volume: | 144 Cu.Ft. |
| Laboratory: | ASG - Port Allen - LA | Cylinder Pressure: | 2015 PSIG |
| Analysis Date: | Mar 14, 2011 | Valve Outlet: | 660 |

**Expiration Date: Mar 14, 2013**

Certification performed in accordance with "EPA Traceability Protocol (Sept. 1997)" using the assay procedures listed. Analytical Methodology does not require correction for analytical interferences. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.
Do Not Use This Cylinder below 150 psig.i.e. 1 Mega Pascal

### ANALYTICAL RESULTS

| Component | Requested Concentration | Actual Concentration | Protocol Method | Total Relative Uncertainty |
|---|---|---|---|---|
| NITRIC OXIDE | 50.00 PPM | 50.06 PPM | G1 | +/- 1% NIST Traceable |
| NITROGEN | Balance | | | |

Total oxides of nitrogen — 50.10 PPM — For Reference Only

### CALIBRATION STANDARDS

| Type | Lot ID | Cylinder No | Concentration | Expiration Date |
|---|---|---|---|---|
| NTRM | 10061119 | cc283757 | 49.73PPM NITRIC OXIDE/NITROGEN | Jul 23, 2016 |

### ANALYTICAL EQUIPMENT

| Instrument/Make/Model | Analytical Principle | Last Multipoint Calibration |
|---|---|---|
| Nicolet 6700 AHR0801556 LNO | FTIR | Feb 28, 2011 |

**Triad Data Available Upon Request**

Notes:

**Approved for Release**

PARR

CAL/Precision